

Quick Start Guide for NanoVi™ Devices

Eng3 Corporation

Setting up the NanoVi™ device

1. Pour distilled, purified, or osmotic water into the glass container, making sure that the water is between the maximum and minimum levels marked on the container.
2. Screw glass container into the humidifier.
3. Insert the humidifier into the fitting on top of the device. The tubes protruding from the humidifier must be properly inserted into the device. There is a small gap between the humidifier and the device when there is a proper connection. Do not try to close this gap.
4. Connect the power cord with the power supply and plug it into the back of the NanoVi™ device. Plug the power cord into an electrical outlet.

Using the NanoVi device

1. The Flex-Arm can be easily adjusted for comfort. If you are using it without a cannula, slide a disposable paper tube onto the end of the arm. It will fit snugly. Adjust the Flex-Arm so the paper tube is two to three inches from your nose.
2. The nasal cannula can be connected to the Flex-Arm. Always use a new nasal cannula for each user to prevent cross-contamination. Position the two air outlets of the cannula into the user's nose and place the thin tubes, like glasses, over the ears.
3. Press the power button on the front of the device. The device automatically starts with a self-test then shows an entry screen.
4. Press the blue start button to begin the session or change the session time by touching the numbers on the screen. Numbers can be cleared with the "C". A session time can be stored in memory as the default time for future sessions by touching the "M".
5. Press the blue start arrow to begin the session. Sessions can be stopped, paused and restarted by touching the appropriate blue control buttons.
6. Use the remote control to adjust the lights illuminating the water. Point the front of the remote control at the base of the water container; the remote must be within a few feet of the glass.
7. The device stops automatically at the end of a session. Press the power button to turn it off. The NanoVi™ device will automatically turn off if it has been left on but not used for 15 minutes.

